

Aquatrak Calibration Certificate

Aquatrak[®] Corporation

1100 E 25th Street

Tel.: 407-814-8680

Fax: 407-814-9744

Sensor S/N:	5002-00198	Calibration Tube Serial No.:	Calibration Date:	Operator
Transducer S/N	2287-AQ0197	A190	September 2, 2010	WS
Controller	R6-00068	F/W Ver.	R1L	


SIGNAL CHARACTERISTICS

1. Ring Period:	2.69	milliseconds
2. Cycle Period:	625.6	microseconds
3. Meas. Clear:	9.67	milliseconds
4. Signal Amplitude @43.00 feet:	372	millivolts

FINAL CALIBRATION

Reference Delta	34.8533 feet		
Measurement 1	Measurement 2	Measured Delta	
42.813	7.960	=	34.853 feet
Difference		±	0.000 feet
Percent Delta		=	0.000 %

OFFSET



OFFSET
0.154 feet
4.679 cm.

Shipped to:

STEVENS