

# Aquatrak Calibration Certificate

Aquatrak<sup>®</sup> Corporation

1100 E 25<sup>th</sup> Street

Tel.: 407-814-8680

Fax: 407-814-9744

Sensor S/N:	5002-00198	Calibration Tube Serial No.:	Calibration Date:	Operator
Transducer S/N	2287-AQ0197	A190	<b>September 2, 2010</b>	WS
Controller	R6-00068	F/W Ver.	R1L	


## SIGNAL CHARACTERISTICS

1. Ring Period:	<b>2.69</b>	milliseconds
2. Cycle Period:	<b>625.6</b>	microseconds
3. Meas. Clear:	<b>9.67</b>	milliseconds
4. Signal Amplitude @43.00 feet:	<b>372</b>	millivolts

## FINAL CALIBRATION

Reference Delta	34.8533 feet	
Measurement 1	Measurement 2	Measured Delta
<b>42.813</b>	<b>7.960</b>	<b>= 34.853 feet</b>
Difference	<b>±</b>	<b>0.000 feet</b>
Percent Delta	<b>=</b>	<b>0.000 %</b>

## OFFSET

	<p>OFFSET</p> <p><b>0.154</b> feet</p> <p><b>4.679</b> cm.</p>
<p>Shipped to:</p> <p><b>STEVENS</b></p>	

AQUATRAK<sup>®</sup> is registered trademark of Aquatrak Corporation